

STATEMENT OF DEFICIENCIES	(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: AC13960105	(X3) DATE SURVEY COMPLETED 07/02/2018
NAME OF PROVIDER OR SUPPLIER TODAY'S WOMEN MEDICAL CENTER	STREET ADDRESS, CITY, STATE, ZIP CODE 3250 S DIXIE HIGHWAY MIAMI, FL 33133	

SUMMARY STATEMENT OF DEFICIENCIES
(FINDINGS PRECEDED BY TAGS AND REGULATORY IDENTIFYING INFORMATION)

0000 - INITIAL COMMENTS

A state re-licensure survey was conducted on . . . , 2018 at Today's Women Medical Center (license #898).

Today's Women Medical Center had deficiencies at the time of the visit.

0156 - Clinic Suppl/eqp-2nd Trimest-Eqpt Maintenance - 59A-9.0225(7), FAC

Based on observation, record review and interview, the facility failed to ensure monitoring equipment and surgical equipment were checked and tested in accordance with manufacturing specifications at designed intervals, not less than annually.

The findings:

Observation on . . . at 10:40 am, revealed various pieces of monitoring equipment and surgical equipment including autoclave machine, suction machine, automatic () monitor, and the defibrillator, with a re-inspection date sticker of . . .

Record review on . . . at 11:00 am, revealed the clinics policy and procedures, states patient monitoring equipment and all surgical instruments shall be checked and/or calibrated and/or tested in accordance with manufacturer's specifications at yearly intervals. Further instructions state: Records shall be kept on each piece of equipment to indicate its history of testing and maintenance.

Interview on . . . at 11:19 am, the Administrator stated the equipment calibration has not expired; the . . . date is the date that the equipment was calibrated. After, looking at the calibration sticker again, the Administrator acknowledged the clinic failed to ensure the monitoring equipment and surgical equipment were checked and tested in accordance with manufacturing specifications at designed intervals, not less than annually.